510(k) SUMMARY

DEC 2 0 2012

Bausch & Lomb PureVision (balafilcon A) Visibility Tinted Contact Lens

1.0 Submitter Information:

Bausch & Lomb
1400 North Goodman Street
Rochester, NY 14609
Contact: Nancy Fehrman
Senior Specialist, Global Regulatory Affairs
1400 North Goodman Street
Rochester, NY 14609
(585) 338-5735 (office)
(585) 338-0702 (fax)
Nancy.a.fehrman@bausch.com
Preparation Date: August 22, 2012

2.0 Device Name:

Trade Name:	Bausch & Lomb PureVision (balafilcon A) Visibility Tinted Contact Lens and Bausch & Lomb PureVision Toric (balafilcon A) Visibility Tinted Contact Lens	
Common Name:	soft (hydrophilic) contact lens	
Device Classification:	Class II (21 CFR 886.5925 (b) (1))	

3.0 Predicate Device

The predicate device, OxyCor (balafilcon A) Visibility Tinted Contact Lens was selected to demonstrate substantial equivalence of the lens material, replacement schedules, and indications for use. OxyCor was cleared in 510(k) Premarket Notification K972454 on August 8, 1997.

4.0 Device Description

The Bausch & Lomb PureVision (balafilcon A) Visibility Tinted Contact Lens and the Bausch & Lomb PureVision Toric (balafilcon A) Visibility Tinted Contact Lens are soft hydrophilic contact lens. The lenses are made from the balafilcon A material, a copolymer of a silicone vinyl carbamate, N-vinyl-pyrrolidone, a siloxane crosslinker and a vinyl alanine wetting monomer, tinted blue with Reactive Blue Dye 246.

5.0 Intended Use

The Bausch & Lomb PureVision (balafilcon A) Visibility Tinted Contact Lens is indicated for the correction of refractive ametropia (myopia and hyperopia) in aphakic and/or not-aphakic persons with non-diseased eyes, exhibiting astigmatism of 2.00 diopters or less, that does not interfere with visual acuity. The lenses may be prescribed for Frequent/Planned Replacement Wear or Disposable Wear in powers ranging from +20.00D to -20.00D for daily wear.

The Bausch & Lomb Pure Vision Toric (balafilcon A) Visibility Tinted Contact Lens is indicated for the correction of refractive ametropia (myopia, hyperopia and astigmatism) in aphakic and/or not-aphakic persons with non-diseased eyes, exhibiting astigmatism of up to 5.00 diopters, that does not interfere with visual acuity. The lens may be prescribed for Frequent/Planned Replacement Wear or Disposable Wear in spherical powers ranging from +20.00D to -20.00D for daily wear.

6.0 Technological Characteristics (comparison to Predicate Device)

The subject devices are identical to the predicate device cleared under K972454. The table below shows a side-by-side comparison of the predicate device to the subject devices:

	OxyCor (balafilcon A) Visibility Tinted	PureVision and PureVision Toric	
Property	Contact Lens	(balafilcon A) Visibility Tinted	
	K972454	Contact Lens	
Material	Balafilcon A		
Water Content	36%		
Physical/optical	Specific Gravity: 1.064]	
properties	Refractive Index: 1.426		
	Surface Character: Hydrophilic		
Oxygen	91 (Boundary and Edge Corrected)		
Permeability*	101 (Boundary and Non-Edge Corrected)	·	
_	*Dk Units = x10 ⁻¹¹		
	[cm³O ₂ (STP)xcm]/(secxcm²xmm Hg)@35°C (Polarographic Method)		
Color Additive	Reactive Blue Dye 246	Same	
Manufacturing	Cook Maria		
Method	Cast Mold		
Lens Design	Spherical and Toric		
Replacement	Frequent/Planned Replacement or		
	Disposable Wear	•	
Base Curve	7.8 to 9.5 mm		
Diameter	13.5 to 15.0 mm		
Sphere Powers	+20.00D to -20.00D		
Cylinder Powers	0.00D to 10.00D		
Axis	0° to 180°		
Center Thickness	0.05 mm to 0.50 mm (varies with power)		
Difference			
	Exhibiting Astigmatism up to 10.00	Exhibiting Astigmatism up to 2.00 diopters (spherical design)	
Indications For Use		or up to 5.00 diopters (toric design)	
	diopters		
036	diopters	, ,	
		(within currently cleared indications)	

7.0 Summary of Non-Clinical Testing

There has been no change made to this device and therefore no non-clinical testing was required to demonstrate substantial equivalence to the predicate device.

8.0 Clinical Testing

There has been no change made to this device and therefore no clinical studies were required to demonstrate the safety or effectiveness of the subject device.

9.0 Substantial Equivalence

There has been no change made to the predicate device. The Indications for Use are within the range previously cleared under K972454. This submission only provides specificity to the Indications for Use for the eye care practitioners regarding prescribing these lenses for frequent/planned replacement wear or disposable wear and to limit the use within the currently cleared indications. Therefore, the Bausch & Lomb PureVision (balafilcon A) Visibility Tinted Contact Lens and the Bausch & Lomb PureVision Toric (balafilcon A) Visibility Tinted Contact Lens are substantially equivalent to the OxyCor (balafilcon A) Visibility Tinted Contact Lens cleared in 510(k) Premarket Notification K972454 on August 8, 1997.





December 20, 2012

Food and Drug Administration 10903 New Hampshire Avenue Document Control Center – WO66-G609 Silver Spring, MD 20993-002

Bausch & Lomb, Incorporated % Ms. Nancy A. Fehrman Senior Specialist, Global Regulatory Affairs 1400 North Goodman Street Rochester, New York 14609

Re: K122575

Trade/Device Name: Bausch & Lomb PureVision (balafilcon A) Visibility Tinted Contact

Lens, Bausch & Lomb Pure Vision Toric (balafilcon A) Visibility

Tinted Contact Lens

Regulation Number: 21 CFR 886.5925

Regulation Name: Soft (hydrophilic) contact lens

Regulatory Class: Class II

Product Code: LPL

Dated: December 7, 2012 Received: December 10, 2012

Dear Ms. Fehrman:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm.

Sincerely yours,

Eric AlMann for

Malvina B. Eydelman, M.D.

Director

Division of Ophthalmic

and Ear, Nose and Throat Devices

Office of Device Evaluation

Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (i	if known):	K122575	
Device Name:	Bausch & Lomb Lens and Bausc Tinted Contact L	ch & Lomb PureVision	A) Visibility Tinted Contact Toric (balafilcon A) Visibility
Indications for U	se:		
indicated for dail in aphakic and/o 2.00 diopters or prescribed for F	y wear for the cor r not-aphakic pers r less, that does requent/Planned	rection of refractive am sons with non-diseased not interfere with vis	sibility Tinted Contact Lens is netropia (myopia and hyperopia) I eyes, exhibiting astigmatism of ual acuity. The lens may be r Disposable Wear in spherical
indicated for dail astigmatism) in astigmatism of u may be prescrib	y wear for the cor aphakic and/or no up to 5.00 diopter oed for Frequent/	rection of refractive am ot-aphakic persons with s, that does not interfe	Visibility Tinted Contact Lens is netropia (myopia, hyperopia and h non-diseased eyes, exhibiting ere with visual acuity. The lens t Wear or Disposable Wear in aily wear.
When prescribe PureVision (bal PureVision Torio disinfected each recommended was recommended was when the prescribe recommended was when prescribe the prescribe was when prescribe was when prescribe was when prescribe was when prescribe was when prescribe was when prescribe was when prescribe was when prescribe was when prescribe when p	afilcon A) Visibi c(balafilcon A) Vis n time it is remo	Planned Replacement Bity Tinted Contact L sibility Tinted Contact L oved from the patient's escribed by the eye can	Wear, the BAUSCH & LOMB ens and BAUSCH & LOMB ens is to be cleaned, rinsed and s eye and discarded after the re practitioner. The lens may be
Visibility Tinted	d for Disposable \ Contact Lens an	Wear, the BAUSCH & IN BAUSCH & LOMB IN be discarded after each	LOMB PureVision (balafilcon A) PureVision Toric (balafilcon A) ch removal.
Prescription Use (Part 21 CFR 80 (PLEASE DO NO	1 Subpart D)	AND/OR THIS LINE-CONTINUE O	Over-The-Counter Use (21 CFR 801 Subpart C) ON ANOTHER PAGE OF NEEDED)
	Concurrence of C	DRH, Office of Device Ev	valuation (ODE)

510(k) Number K122575

Division of Ophthalmic, Neurological and Ear,

Nose and Throat Devices